http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 11-Mar-2014 UK32S11123558-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
EN ISO 9001:2008 Quality Management System
EN ISO 13485:2012 Quality System - Medical Devices

Sampling Systems Ltd
4 & 5 Forge Mills Park
Station Road
Coleshill
Warwickshire B46 1JH
UNITED KINGDOM

Order Information

Account Number:	100494
Synergy Health Sales Part Reference:	1060501
Customer Reference Number:	S14-15351
Product Description:	8015NR-170S RED NARROW M/SP 25-45KGY
Validation Reference:	4.1529
Quantity Received:	50
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	LOT 2; 1 PLT; MON 03.03.14; B/NO. 25854

Irradiation Data

Date and Time of Irradiation:	11-Mar-2014 00:56
Reference Dose Range kGy:	35.3 - 35.3
Calculated Minimum Dose kGy:	30.2
Calculated Maximum Dose kGy:	37.9

The Expiry Date is 2 years from the Date of Irradiation provided the bag remains intact.
Expiry Date text added to Certificate by Sampling Systems Ltd.

Irradiation Release Authorised By Synergy Health plc