

<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 03-Nov-2016

UK32S11712912-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products
EN ISO 9001:2008 Quality Management System
EN ISO 13485:2012 Quality System - Medical Devices

Sampling Systems Ltd
4 & 5 Forge Mills Park
Station Road
Coleshill
Warwickshire B46 1JH
UNITED KINGDOM

Order Information

Account Number:	100494
Synergy Health Sales Part Reference:	1002970
Customer Reference Number:	S16-18082
Product Description:	8640H-01S VOL SPOON 1ML 25-45kGy
Validation Reference:	4.1529
Quantity Received:	59
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	B/NO 47372 WED 21ST SEP 2016 1 PLT LOT 4

Irradiation Data

Date and Time of Irradiation:	03-Nov-2016 10:22
Reference Dose Range kGy:	35.4 - 35.4
Calculated Minimum Dose kGy:	30.3
Calculated Maximum Dose kGy:	38.0

Expiry Date

The Expiry Date is printed on the box label and on the bag label

Expiry Date text added to Certificate by Sampling Systems Ltd.

Irradiation Release Authorised By Synergy Health plc

Processing Site: Roysdale Way, Euroway Trading Estate, Bradford, BD4 6SE Phone No: +44 (0) 1274 686011

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069