

1. Part Numbers: 5035S-600

2. Photograph:



3. Product Information:

| | | | | | | | | | | | | | | | |
|---------------------------------|--|-----|----------------------|------------|----------------------|-----------|----------------------|--------|----------------------|--------|----------------------|------|---|--------|-----------------|
| Description | Liquid Master 600mm – Stainless Steel | | | | | | | | | | | | | | |
| Materials of Construction: | <table border="0"> <tr> <td>Tip</td> <td>316L Stainless Steel</td> </tr> <tr> <td>Outer Tube</td> <td>316L Stainless Steel</td> </tr> <tr> <td>Inner Rod</td> <td>316L Stainless Steel</td> </tr> <tr> <td>Piston</td> <td>316L Stainless Steel</td> </tr> <tr> <td>Handle</td> <td>316L Stainless Steel</td> </tr> <tr> <td>Seal</td> <td>Silicone (conforms to FDA CFR 177.2600)</td> </tr> <tr> <td>Spring</td> <td>Stainless Steel</td> </tr> </table> | Tip | 316L Stainless Steel | Outer Tube | 316L Stainless Steel | Inner Rod | 316L Stainless Steel | Piston | 316L Stainless Steel | Handle | 316L Stainless Steel | Seal | Silicone (conforms to FDA CFR 177.2600) | Spring | Stainless Steel |
| Tip | 316L Stainless Steel | | | | | | | | | | | | | | |
| Outer Tube | 316L Stainless Steel | | | | | | | | | | | | | | |
| Inner Rod | 316L Stainless Steel | | | | | | | | | | | | | | |
| Piston | 316L Stainless Steel | | | | | | | | | | | | | | |
| Handle | 316L Stainless Steel | | | | | | | | | | | | | | |
| Seal | Silicone (conforms to FDA CFR 177.2600) | | | | | | | | | | | | | | |
| Spring | Stainless Steel | | | | | | | | | | | | | | |
| Method of Construction: | All components screw together | | | | | | | | | | | | | | |
| External Surface Finish: | Better than 0.5 microns Ra | | | | | | | | | | | | | | |
| Nominal Length: | 600mm | | | | | | | | | | | | | | |
| Nominal Weight: | 1.4kg | | | | | | | | | | | | | | |
| Diameter of Outer Tube: | 25mm | | | | | | | | | | | | | | |
| Diameter of Hole in Tip: | 6mm | | | | | | | | | | | | | | |
| Recommended Storage Conditions: | Dry and ambient temperature | | | | | | | | | | | | | | |
| Cleaning: | This product should be cleaned before first used | | | | | | | | | | | | | | |

4. Approval:

Signed:



Name: Kuldeep Flora

Position: Quality Manager

Date: 13-Oct-20