

## **Specification Sheet**

Specification Number: 8027A-190
Revision: 01
Date Issued: 17-Mar-25

**1. Part Number:** 8027A-190

## 2. Photograph:



## 3. Product Information:

Description: SteriWare® Laboratory Knife 190mm
Use: Single use, disposable laboratory knife

Material of Construction: High Impact Polystyrene (HIPS) – virgin

• Conforms to FDA CFR 177.1640

• Conforms to EU Regulations 10/2011

• Conforms to EC Regulations 1935/2004

• BSE/TSE free

Method of Construction: Single piece, injection moulded

Overall Length: 190mm
Handle Length: 125mm
Maximum Height: 16mm
Nominal Weight: 4g

Moulding Environment: Class 100,000 Cleanroom
Primary Packing Environment: Class 10,000 Cleanroom

Individually Bagged? Yes (heat sealed PE/PET bag)

Method of Sterilisation: N/A

Quantity per Box: 100

Recommended Storage Conditions: Dry and ambient temperature

Shelf Life: 5 years from the date of manufacture
ATEX: This product has not been ATEX tested





## **Specification Sheet**

Specification Number: 8027A-190
Revision: 01
Date Issued: 17-Mar-25

**4. Change Control:** Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the

SteriWare range please email your contact details to

change.control@sampling.com

**5. Country of Origin:** Made and packed in the UK.

**6. Repeated Use:** This is a single-use, disposable device. Repeated use will lead to

failure.

7. Approval: Signed: Name: Kuldeep Flora

Position: Quality Manager

Date: 17-Mar-25