

http://www.synergyhealthplc.com

## **Certificate of Irradiation**

Date Issued: 12-Apr-2013 UK32S11012314-2-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

ISO 11137:2006 Sterilisation of Health Care Products ISO 9001:2008 Quality Management System ISO 13485:2012 Quality System - Medical Devices

Sampling Systems Ltd 4 & 5 Forge Mills Park Station Road Coleshill Warwickshire B46 1JH UNITED KINGDOM

## **Order Information**

Customer Reference Number:	S13-14378
Product Description:	8050H-1000S DIS LIQ 25-45kGy
Validation Reference:	4.1052
Quantity Received:	20
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	B/No.17976 , 14 PLTS TOTAL ORDER TUES 2.4.13
Shop Order Number:	S11012314-2-1
Synergy Health Sales Part Reference:	1002933

Date and Time of Irradiation:	12-Apr-2013 11:06
Reference Dose Range kGy:	34.9 - 36.4
Calculated Minimum Dose kGy:	30.2
Calculated Maximum Dose kGy:	38.4

Irradiation Data

The Expiry Date is 2 years from the Date of Irradiation provided the bag remains intact.

Expiry Date text added to Certificate by Sampling Systems Ltd.

Irradiation Release Authorised By Synergy Health plc