

http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 04-Sep-2013 UK32S11048051-4-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Sampling Systems Ltd 4 & 5 Forge Mills Park **Station Road** Coleshill Warwickshire B46 1JH **UNITED KINGDOM**

Order Information

| Account Number: | 100494 | |
|--------------------------------------|---|--|
| Synergy Health Sales Part Reference: | 1003504 | |
| Customer Reference Number: | S13 - 14764 | |
| Product Description: | 5055G- 50TS GLASS THUMB TUBES 25-45kGy | |
| Validation Reference: | 4.1072 | |
| Quantity Received: | 5 | |
| Customer Minimum Specification kGy: | 25.0 | |
| Customer Maximum Specification kGy: | 45.0 | |
| Customer Unit Lot/Batch Number: | B/No. S08 9941 , 5 PLTS TOTAL ORDER THURS 22.8.13 | |

| Irradiation Data | | |
|-------------------------------|-------------------|--|
| Date and Time of Irradiation: | 04-Sep-2013 11:31 | |
| Reference Dose Range kGy: | 34.3 - 34.3 | |
| Calculated Minimum Dose kGy: | 29.9 | |
| Calculated Maximum Dose kGy: | 36.1 | |

The Expiry Date is 2 years from the Date of Irradiation provided the bag remains intact.

Expiry Date text added to Certificate by Sampling Systems Ltd.

Irradiation Release Authorised By Synergy Health plc