

http://www.synergyhealthplc.com

Certificate of Irradiation

UK32S11083311-1-1 Date Issued: 06-Dec-2013

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products

EN ISO 9001:2008 Quality Management System

EN ISO 13485:2012 Quality System - Medical Devices

Sampling Systems Ltd 4 & 5 Forge Mills Park **Station Road** Coleshill Warwickshire B46 1JH **UNITED KINGDOM**

Order Information

Account Number:	100494
Synergy Health Sales Part Reference:	1002927
Customer Reference Number:	S13-15102
Product Description:	8045A-30S DISP PHARMA SCOOP 25-45kGy
Validation Reference:	4.0511
Quantity Received:	74
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	LOT 1; TOTAL ORDER 15 PLTS; WED 04.12.13; B/NC 22143

irradiation Data	
Date and Time of Irradiation:	06-Dec-2013 13:27
Reference Dose Range kGy:	35.3 - 36.2
Calculated Minimum Dose kGy:	29.7
Calculated Maximum Dose kGy:	38.7

Irradiation Data

The Expiry Date is 2 years from the Date of Irradiation provided the bag remains intact.

Expiry Date text added to Certificate by Sampling Systems Ltd.

Irradiation Release Authorised By Synergy Health plc