



<http://www.synergyhealthplc.com>

# Certificate of Irradiation

Date Issued: 14-Jul-2014

UK32S11192124-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products  
EN ISO 9001:2008 Quality Management System  
EN ISO 13485:2012 Quality System - Medical Devices

Sampling Systems Ltd  
4 & 5 Forge Mills Park  
Station Road  
Coleshill  
Warwickshire B46 1JH  
UNITED KINGDOM

---

## Order Information

Account Number:	100494
Synergy Health Sales Part Reference:	1074169
Customer Reference Number:	S14-15705
Product Description:	8045R-250S RED DIS PHARMA 25-45KGY
Quantity Received:	22
Customer Unit Lot/Batch Number:	B/NO: 28707 fri 27.06.14 2plts lot 3
Other Process Details:	Released against validation 4.1742. Calculated minimum dose = 29.3kGy. Calculated maximum dose = 36.3kGy.

---

## Irradiation Data

Date and Time of Irradiation:	13-Jul-2014 05:46
-------------------------------	-------------------

### Expiry Date

The Expiry Date is printed on the box label and on the bag label

Expiry Date text added to Certificate by Sampling Systems Ltd.

---

Irradiation Release Authorised By Synergy Health plc

---

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM  
Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069