



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 29-Jul-2014

UK32S11208492-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

Sampling Systems Ltd
 4 & 5 Forge Mills Park
 Station Road
 Coleshill
 Warwickshire B46 1JH
 UNITED KINGDOM

Order Information

Account Number:	100494
Synergy Health Sales Part Reference:	1002961
Customer Reference Number:	S14-15780
Product Description:	8088A-10S SAMPLE SPOON 10ML 25-45kGy
Validation Reference:	4.0890
Quantity Received:	60
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	B/NO.24053 MON 28.07.14 1 PLT LOT 2

Irradiation Data

Date and Time of Irradiation:	29-Jul-2014 07:17
Reference Dose Range kGy:	34.3 - 34.9
Calculated Minimum Dose kGy:	29.6
Calculated Maximum Dose kGy:	36.4

Expiry Date

The Expiry Date is printed on the box label and on the bag label

Expiry Date text added to Certificate by Sampling Systems Ltd.

Irradiation Release Authorised By Synergy Health plc

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
 Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069