



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 05-Nov-2014

UK32S11270155-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
EN ISO 9001:2008 Quality Management System
EN ISO 13485:2012 Quality System - Medical Devices

Sampling Systems Ltd
4 & 5 Forge Mills Park
Station Road
Coleshill
Warwickshire B46 1JH
UNITED KINGDOM

Order Information

Account Number:	100494
Synergy Health Sales Part Reference:	1075054
Customer Reference Number:	S14-16010
Product Description:	8081P-3000S 25-45KGY
Validation Reference:	4.1732
Quantity Received:	32
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	B/N 31250 MON 03.11.14 4 PLTS LOT 6

Irradiation Data

Date and Time of Irradiation:	05-Nov-2014 07:11
Reference Dose Range kGy:	38.0 - 38.7
Calculated Minimum Dose kGy:	31.3
Calculated Maximum Dose kGy:	41.0

Expiry Date

The Expiry Date is printed on the box label and on the bag label

Expiry Date text added to Certificate by Sampling Systems Ltd.

Irradiation Release Authorised By Synergy Health plc

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069