

http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 06-Sep-2015 UK32S11455910-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Sampling Systems Ltd 4 & 5 Forge Mills Park Station Road Coleshill Warwickshire B46 1JH UNITED KINGDOM

Order Information

Account Number:	100494
Synergy Health Sales Part Reference:	1003446
Customer Reference Number:	S15-16873 & S15-16882
Product Description:	8045R-1000S STERILE SCOOPS 25-45kGy
Validation Reference:	4.1025
Quantity Received:	192
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	B/N 37514 THUR 27 AUG 2015 10 PLTS LOT 2

Irradiation Data	
Date and Time of Irradiation:	06-Sep-2015 18:26
Reference Dose Range kGy:	34.3 - 36.2
Calculated Minimum Dose kGy:	29.4
Calculated Maximum Dose kGv	36.2

Expiry Date

The Expiry Date is printed on the box label and on the bag label Expiry Date text added to Certificate by Sampling Systems Ltd.

Irradiation Release Authorised By Synergy Health plc