

http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 20-Jan-2016 UK32S11545602-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Sampling Systems Ltd 4 & 5 Forge Mills Park Station Road Coleshill Warwickshire B46 1JH UNITED KINGDOM

Order Information

Account Number:	100494	
Synergy Health Sales Part Reference:	1002946	
Customer Reference Number:	S16-17285	
Product Description:	8060H-501S DISP POWDER 25-45kGy	
Validation Reference:	4.2001	
Quantity Received:	20	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	45.0	
Customer Unit Lot/Batch Number:	BATCH NO 41981 FRI 15TH JAN 2016 1 PLT LOT 3	

Irradiation Data		
Date and Time of Irradiation:	20-Jan-2016 10:39	
Reference Dose Range kGy:	36.3 - 36.3	
Calculated Minimum Dose kGy:	30.8	
Calculated Maximum Dose kGy:	38.8	

Expiry Date

The Expiry Date is printed on the box label and on the bag label Expiry Date text added to Certificate by Sampling Systems Ltd.

Irradiation Release Authorised By Synergy Health plc