

http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 23-Jan-2016 UK32S11546309-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Sampling Systems Ltd 4 & 5 Forge Mills Park Station Road Coleshill Warwickshire B46 1JH UNITED KINGDOM

Order Information

Account Number:	100494
Synergy Health Sales Part Reference:	1006380
Customer Reference Number:	S16-17285
Product Description:	8045A-500S DISP PHARMA SCOOP 25-45KGY
Validation Reference:	4.1544
Quantity Received:	90
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	BATCH NO: 42008 FRI 15TH JAN 2016 5 PLTS LOT 1

Date and Time of Irradiation:	23-Jan-2016 17:00
Reference Dose Range kGy:	35.6 - 35.6
Calculated Minimum Dose kGy:	29.2
Calculated Maximum Dose kGy:	36.2

Expiry Date

The Expiry Date is printed on the box label and on the bag label

Expiry Date text added to Certificate by Sampling Systems Ltd.

Irradiation Release Authorised By Synergy Health plc