

http://www.synergyhealthplc.com

## **Certificate of Irradiation**

Date Issued: 21-0ct-2016 UK32S11698507-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Sampling Systems Ltd 4 & 5 Forge Mills Park Station Road Coleshill Warwickshire B46 1JH UNITED KINGDOM

## **Order Information**

Account Number:	100494
Synergy Health Sales Part Reference:	1002924
Customer Reference Number:	<b>S16-17997</b>
Product Description:	8045A-1000S STERILE SCOOPS 25-45kGy
Validation Reference:	4.1025
Quantity Received:	180
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	B/NO 47287 WED 31ST AUG 2016 10 PLTS LOT 2

Date and Time of Irradiation:	21-0ct-2016 02:55
Reference Dose Range kGy:	34.2 - 36.3
Calculated Minimum Dose kGy:	29.3
Calculated Maximum Dose kGy:	36.3

Irradiation Data

## **Expiry Date**

The Expiry Date is printed on the box label and on the bag label

Expiry Date text added to Certificate by Sampling Systems Ltd.

Irradiation Release Authorised By Synergy Health plc