http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 03-Feb-2017 UK32S11794790-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products
EN ISO 9001:2008 Quality Management System
EN ISO 13485:2012 Quality System - Medical Devices

Sampling Systems Ltd
4 & 5 Forge Mills Park
Station Road
Coleshill
Warwickshire B46 1JH
UNITED KINGDOM

Order Information

Account Number:	100494
Synergy Health Sales Part Reference:	1100698
Customer Reference Number:	S17-18465
Product Description:	8640H-0264S VOLMETRIC SPOON 25-45KGY
Validation Reference:	4.1331
Quantity Received:	10
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	B/NO 49208 FRI 20TH JAN 2017 1 PLT LOT 7

Irradiation Data

Date and Time of Irradiation:	03-Feb-2017 06:39
Reference Dose Range kGy:	33.8 - 33.8
Calculated Minimum Dose kGy:	28.4
Calculated Maximum Dose kGy:	35.5

Expiry Date
The Expiry Date is printed on the box label and on the bag label
Expiry Date text added to Certificate by Sampling Systems Ltd.

Irradiation Release Authorised By Synergy Health plc